

Certificate of Compliance

Certificate No.

: BSTXD221222298801EC

Applicant

Ziglite Smart Health Care Products Co., Ltd

Block B3,18/F, Bonsun Industrial Building, 364-366 Sha Tsui

Road, Tsuen Wan, H.K.

Manufacturer

Ziglite Smart Health Care Products Co., Ltd

Block B3,18/F, Bonsun Industrial Building, 364-366 Sha Tsui

Road, Tsuen Wan, H.K.

Product Name

: Hand Strectching Rehabilition Therapy Device

Trademark

7 IGLITE

Main Test Model

: SHC

Additional Model

: SBA2, SHA, SKE

Test Standard

: EN IEC 55014-1:2021

EN IEC 61000-3-2:2019/A1:2021

EN 61000-3-3:2013/A2:2021/AC:2022-01

EN IEC 55014-2:2021

As shown in the Test Report No.

BSTXD221222298801ER

The EUT described above has been tested by us with the listed standards and found in compliance with the council EMC directive 2014/30/EU. It is possible to use CE marking to demonstrate the compliance with this EMC Directive.

The certificate applies to the tested sample above mentioned only and shall not imply an assessment of the whole production.



CE

Christina Dens Manager Dec. 22, 2022

BST Testing (Shenzhen) Co.,Ltd.

Add: No.7, New Era Industrial Zone, Guantian, Bao'an District, Shenzhen, Guangdong, China

Certificate Search: http://www.bst-lab.com, Tel:400-962-6168, 8009990305, E-mail:christina@bst-lab.com